

Firmaet
Uson Plast A/S
Finlandsvej 9

5700 Svendborg

att: Else Lindhard

RESULT FOR EO STERILIZATION

The following products, with the belowmentioned sterilization lot numbers have been sterilized by means of Ethylene Oxide under the specification for Programme S01:

Test Methods

Results

DS/EN ISO 11138-1:2006 Biological Indicators for sterilization

passed

Recorded process parameters have been found to be within the sterilization programme specifications.

The validation of the sterilization process was performed according to:

DS/EN ISO 11135-1:2007

Sterilization of medical devices – validation and routine control of Ethylene Oxide sterilization.

EN 556

Sterilization of medical devices – requirements for medical devices to be labelled "STERILE".

Ster.Batch No.	Lot No.	Pallets
42130318		5

date: 21.03.2013

Certified by:
CODAN Medical ApS


Jytte Hjelmsmark
Q.A. Dept.